

# DECLARATION OF CONFORMITY

**We:**

**Hanil Science Industrial Co., Ltd.**

#916-5, Daejeon-dong, Gangneung-si, Gangwon-do, Republic of Korea.

DECLARE UNDER OUR SOLE RESPONSIBILITY THAT THE PRODUCT:

**PRODUCT :**

**Medical Centrifuge**

**TYPE-DESIGNATION :**

**FLETA 40, FLETA 40P**

TO WHICH THIS DECLARATION RELATES IS IN CONFORMITY WITH THE FOLLOWING STANDARD(S) OR OTHER NORMATIVE DOCUMENT(S)

**EMCD: EN 55011:2009+A1:2010 [Group1, Class A Equipment]  
EN 60601-1-2:2007  
EN 61000-3-2:2006+A1:2009+A2:2009  
EN 61000-3-3:2008**


**LVD: EN 60601-1:2006**

FOLLOWING THE PROVISIONS OF THE

**LOW VOLTAGE DIRECTIVE 2006/95/EC,  
Medical Device Directive 93/42/EC,  
and the EMC Directive 2004/108/EC.**

**March 6, 2014**

(Date of issue)



**/President**

(Name and signature of authorized person)





# EMC TEST REPORT

Medical Electrical Equipment

Emission       Immunity

Test Report File No.	13-IST-0876	<input checked="" type="checkbox"/> Basic	<input type="checkbox"/> Alternate
Date of Receipt	Oct. 15, 2013	Begin of test date	Dec. 17, 2013
Date of Issue	Jan. 02, 2014	End of test date	Dec. 23, 2013

Kind of Product	Medical Centrifuge
Basic Model(s)	FLETA 40
Multi Model(s)	-

Applicant/ Manufacturer	Hanil Science Industrial Co., Ltd.
Address	#916-5, Daejeon-dong, Gangneung-si, Gangwon-do, Republic of Korea

Standard(2004/108/EC)	EN 55011:2009 +A1:2010 [Group1, Class A Equipment]
Standard(93/42/EEC)	EN 60601-1-2:2007 EN 61000-3-2:2006 +A1:2009, +A2:2009, EN 61000-3-3:2008

**Test Result**       Positive       Negative

Tested By

Reviewed By

H.K.LEE

S.J.CHO

**Comment (s)**

- This test report consists of 41 pages with appendix.
- The test result only responds to the tested sample. This report shall not be reproduced except in full, without the allowance of the test laboratory.

